HAC-RF Emission

Communication System: UID 0, CW; Frequency: 835 MHz; Duty Cycle: 1:1

Phantom section: RF Section

DASY5 Configuration:

- Probe: ER3DV6 - SN2509; ConvF(1, 1, 1); Calibrated: 5/14/2015;

- Sensor-Surface: (Fix Surface)

- Electronics: DAE4 Sn1259; Calibrated: 1/14/2015

- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BB

- Measurement SW: DASY52, Version 52.8 (7); SEMCAD X Version 14.6.10 (7164)

Dipole E-Field measurement/835 MHz/Hearing Aid Compatibility Test at 15mm distance

Date: 1/8/2016

(41x361x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 124.2 V/m; Power Drift = 0.06 dB


PMR not calibrated. PMF = 1.000 is applied.

E-field emissions = 118.3 V/m

Near-field category: M4 (AWF 0 dB)

PMF scaled E-field

Grid 1 M4	Grid 2 M4	Grid 3 M4
113.6 V/m	118.3 V/m	117.2 V/m
Grid 4 M4	Grid 5 M4	Grid 6 M4
64.00 V/m	66.04 V/m	65.65 V/m
Grid 7 M4	Grid 8 M4	Grid 9 M4
102 2 37/	102 7 37/	102.8 V/m



0 dB = 118.3 V/m = 41.46 dBV/m

HAC-RF Emission

Communication System: UID 0, CW; Frequency: 1880 MHz; Duty Cycle: 1:1

Phantom section: RF Section

DASY5 Configuration:

- Probe: ER3DV6 - SN2509; ConvF(1, 1, 1); Calibrated: 5/14/2015;

- Sensor-Surface: (Fix Surface)

- Electronics: DAE4 Sn1259; Calibrated: 1/14/2015

- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BB

- Measurement SW: DASY52, Version 52.8 (7); SEMCAD X Version 14.6.10 (7164)

Dipole E-Field measurement/1880 MHz/Hearing Aid Compatibility Test at 15mm

distance (41x181x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 129.5 V/m; Power Drift = 0.05 dB


PMR not calibrated. PMF = 1.000 is applied.

E-field emissions = 90.03 V/m

Near-field category: M3 (AWF 0 dB)

PMF scaled E-field

Grid 1 M3	Grid 2 M3	Grid 3 M3
88.11 V/m	90.03 V/m	88.47 V/m
Grid 4 M3	Grid 5 M3	Grid 6 M3
68.58 V/m	70.73 V/m	70.56 V/m
Grid 7 M3	Grid 8 M3	Grid 9 M3
85.12 V/m	87.47 V/m	86.56 V/m



0 dB = 90.03 V/m = 39.09 dBV/m

Date: 1/8/2016