




## Product Similarity Declaration

**Date:** Nov. 16, 2023

**Model Number:** KD-5811BT, KD-5810BT, KD-5920BT

To Whom It May Concern,

We, ANDON HEALTH CO., LTD., declare that EUT Devices Model KD-5811BT, KD-5810BT, KD-5920BT have some minor differences in the model name, appearance, software version and power supply, but are the same on the hardware version, bluetooth version, PCB layout, and circuit schematic.

Model name	KD-5811BT	KD-5810BT	KD-5920BT
Appearance			
Software version	BPM-SZC900-V0028-1.0.0	BPM-SZC900-V0027-1.0.0	BPM-SZC900-V0038-1.0.0
Power supply	Medical AC adapter: 5V 1A Battery: 4 ×1.5V SIZE AA	Medical AC adapter: 5V 1A Battery: 4 ×1.5V SIZE AA	Medical AC adapter: NA Battery: 4 ×1.5V SIZE AAA

Should you have any questions or comments regarding this matter, please have my best attention.

Sincerely,

(Signed)

A handwritten signature in black ink that reads "Ming Cong". The signature is written in a cursive, flowing style.

Name / Title: Ming Cong / PM

Company: Andon Health Co., Ltd

Address: No.3 JinPing Road, Nankai District, Tianjin, China 300190