

RF Exposure Statement: 15045157 001

Page 1 of 1

Client: Andon Health Co., Ltd.
 No.31 Changjiang Road, Nankai District Tianjin 300193, P.R. China

Test item: Automatic Blood Pressure Monitor

Identification: KD-923

FCC Requirement

According to KDB 447498, SAR evaluation specified in FCC 2.1093 is not required for portable equipment if the transmitter power is below the following threshold:

Tunable Range		Center of Tunable Band [GHz]	60/f SAR Limitation based on Center of Band [mW]
Lowest Frequency [GHz]	Highest Frequency [GHz]		
2.402	2.480	2.441	24.58

Measurement Result

The maximum measured transmitter power is the following:

Conducted Output Power P_{out} [mW]	Maximum Antenna Gain [dBi]	P_{out} EIRP [mW]
7.533	2.28	12.74

Conclusion

SAR evaluation is not required since the maximum transmitter power P_{out} (both conducted and e.i.r.p.) is below the FCC threshold.

Refer to test report 15045157 001 for more details.