

Berne, 13 Feb. 2013

Federal Communication Commission
Equipment Authorization Division, Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

Certification and Engineering Bureau
Industry Canada
Spectrum Engineering Branch
3701 Carling Avenue, Building 94
Ottawa, Ontario K2H 8S2

Declaration on Electrically Identical Models

TO WHOM IT MAY CONCERN

We, Sonic Innovations, Inc., declare on our sole responsibility that the new models

BL100 ITCPDW, BL80 ITCPDW, BL100 ITEPDW, BL80 ITEPDW with
FCC ID: **ZTOF2ITE** and IC: **9799A-F2ITE01**

are identical in hardware and software to the certified model

ITC/ITE Fusion 2 under FCC ID: **U28FU2ITE** and IC: **1350B-FU2ITE**

The products are hearing instruments. The only differences between the new models and the original model are:

- Models BL100 xxx and BL80 xxx offer different audiological signal processing. This is implemented by different parameter sets for the digital signal processing of the audio path.
- Models BLnnn ITCPDW and BLnnn ITEPDW offer different electro-acoustical performance (maximum possible sound pressure level). This is implemented by the use of different electro-acoustic transducers (loudspeaker).

We attest that above changes are not relevant for any RF behavior subject to regulatory items. Therefore all test reports remains applicable.

If you have any questions, please feel free to contact us at the address shown below

Sincerely,



(signature)
David Scott, Quality Coordinator
Sonic Innovations Inc.
2501 Cottontail Land
Somerset, NJ 08873
USA
Phone: +41 31 560 21 25

