

Measurement Report for Device
Device Under Test Properties

Model, Manufacturer	Dimensions [mm]	Software Version	DUT Type
Device,	70.0 x 50.0 x 8.0	3.2.0.1840	5G Verification Source

Exposure Conditions

Phantom Section	Position, Test Distance [mm]	Frequency [MHz]	Conversion Factor
5G	Front, 10.00	10000.0	1.0

Hardware Setup

Phantom	Medium	Probe, Calibration Date	DAE, Calibration Date
mmWave - 1044	Air -	EUmmWV3 - SN9424_F1-55GHz, 2024-03-12	DAE4 Sn1696, 2023-10-23

Scans Setup

Scan Type	5G Scan
Grid Extents [mm]	120.0 x 120.0
Grid Steps [lambda]	0.25 x 0.25
Sensor Surface [mm]	10.0

Measurement Results

Date	2024-07-18
Avg. Area [cm ²]	4.00
psPDn+ [W/m ²]	61.4
psPDtot+ [W/m ²]	61.6
H _{max} [A/m]	0.436
E _{max} [V/m]	164
max _(stot) [W/m ²]	69.3
Power Drift [dB]	0.10
IPDn	112

