

DECLARATION OF SIMILARITY

March 24, 2016

To:
Bay Area Compliance Laboratories Corp.
1274 Anvilwood Ave.
Sunnyvale, CA 94089
Phone: 408-732-9162, Fax: 408-732-9164
<http://www.baclcorp.com>

Dear Sir or Madam:

We *Siemens Medical Solutions USA Inc.* hereby declare that product: *Handheld Probe*, model(s): *L8-3, C5-2* is/are electrically identical with the same electromagnetic emissions and electromagnetic compatibility characteristics as model: *L13-5* tested by BACL, the results of which are featured in BACL project: *R1602221*

A description of the differences between the tested model and those that are declared similar are as follows:

The ACUSON Freestyle handheld probes, Models L8-3, L13-5 and C5-2 are all equivalent electrically and use the same internal circuitry. The schematics and materials for all three models are identical. The difference is that each model uses a different ultrasonic transducer array (the nosepiece), which is driven by the internal circuitry. The different nosepieces are optimized for different procedures within diagnostic ultrasound.

Please contact me should there be need for any additional clarification or information.

Best Regards,



Larry Engle
Director, Quality
Phone: 610.834.1220 x102
Email: larry.engele@siemens.com
Siemens Medical Solutions USA Inc.
5168 Campus Drive, Plymouth Meeting, PA 19462, USA